



Boaray600 Anesthesia Workstation

- 20-1500ml (VCV); Applicable to adult, pediatric, infant
- Various modes for different needs of all patients
- Anesthesia Agent, CO₂, and SpO₂ monitoring modules (optional)
- 15 inch TFT touch screen & big font view for easy observation
- Touch screen & navigation knob for simple operation
- Simultaneous display of 4 waveforms and 2 loops
- 6 tubes of flowmeters for O₂, N₂O, Air
- Highly integrated absorber, autoclavable in 134°C to avoid cross-infection (optional)
- The absorber could be automatically heated, to avoid condensation water
- Electronic PEEP
- Advanced proportional solenoid technology to realize accurate ventilation control
- Active AGSS system deal with exhaust gas and reduce pollution, providing more safety protection (optional)

Specification	Boaray600
Ventilation modes:	VCV,PCV,SIMV(V)+PS,SIMV(P)+PS,Sigh,Spont/CPAP,PRVC(optional), Cardio Pulmonary Bypass,Standby,Manual
Tidal Volume:	20-1500 mL
Ventilation frequency:	1-100 bpm
I:E:	4:1-1:10 (increment:0.5)
Ti:	0.1-10S (increment:0.1S)
Inspiratory pause (Tp):	OFF, 5%-50%
Flow Trigger:	1-15L/Min
Pressure control:	5-70 cm H ₂ O
Pressure support:	5-60 cm H ₂ O
Pressure limited:	PEEP+5-PEEP+70 cm H ₂ O
Electronic PEEP:	OFF, 4-30 cm H ₂ O
Monitoring:	Volume: V _{TI} , V _{TE} , MV, MVspont Pressure: Ppeak, PEEP, Pplat, Pmean, Pmin Freq, Fspont, I:E, Compliance, Resistance, FiO ₂ SpO ₂ Module (optional): SpO _{2r} , PR CO ₂ Module (optional): EtCO _{2r} , FiCO ₂ Anesthetic Agent Module (optional): Anesthetic agent, N ₂ O, CO ₂ and MAC _{Age}
Waveform:	P-T, F-T, V-T, Optional: SpO ₂ -T, CO ₂ -T, FiO ₂ -T, Anesthetic agent-T
Loop:	P-V, F-V, F-P, Optional: Volume-CO ₂
Trend:	Ppeak, FiO _{2r} , MV, PEEP, V _{TE}
Alarm:	Audio and visual alarms with three levels Volume: MV high/low, V _{TE} high/low Pressure: Paw high/low, Negative pressure, Apnea Continuous high airway pressure, Low O ₂ supply pressure Freq high/low, FiO ₂ high/low, EtCO ₂ high/low (optional), SpO ₂ low (optional) Power: AC power failure, Low battery power, Exhausted battery failure

Vaporizer	
2 stations with interlock (Selectatec compatible)	
Agent type:	Halothane, Enflurane, Isoflurane, Sevoflurane
Filler type:	Pour fill, Quik fil [®] (only Sevoflurane)

Pneumatic Specification	
Gas supply:	O ₂ , N ₂ O, Air: 0.28-0.6 Mpa
Flowmeter:	O ₂ /N ₂ O/Air: 0-1L, 1-10L/Min
Gas system:	O ₂ deficiency alarm, Hypoxic guard system, O ₂ flush: 35-75L/Min
Working mode:	Closed, Semi-closed, Semi-open, Open
Driven mode:	Pneumatically driven and electronically controlled
Operating mode:	Manual/Vent; APL: 2-70 cmH ₂ O
ACGO Connector:	22mm

Physical Specification	
Dimensions (H×W×D):	1550*1000* 870 mm
Communication Interface:	VGA, RS232
Screen:	15 inch TFT touch screen
Net weight:	108 kg
Power:	100-240VAC, 50/60Hz, 6.5A(Max)
Lithium Battery:	Minimum 120 minutes

Shenzhen PRUNUS Medical Co., Ltd.

9th Floor, Prunus Building, Block C, Funing High-Tech Industrial Zone,
No.71 Xintian Road, Baoan District, Shenzhen, Guangdong,
P.R.China

Tel: +86 755 2689 9781
Fax: +86 755 2689 9789

E-mail: international@prunusmedical.com
Web: www.prunusmedical.com